



Select the assay(s) for which training was given:

SARS Antigen

Influenza A+B

Flu + SARS Antigen

RSV

Strep A+

Fill in or click the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the assay(s) selected above. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current Package Insert instructions and/or Quick Reference instructions on quidel.com.

1. How often should the Calibration Check procedure be run?

- a. Every 14 days
- b. Every 30 days
- c. Anytime
- d. With every new kit lot

2. How often does the manufacturer recommend that external controls be run on the kit?

- a. Once every 20 tests
- b. Once per each shipment of kits
- c. Every first sample tested each day
- d. Once per each untrained operator, once for each new shipment of kits and/or new lot of kits

3. At what temperature do the kits need to be stored?

- a. 2°C to 8°C (refrigerated)
- b. 15°C to 30°C (room temperature)
- c. 2°C to 30°C (room temperature or refrigerated)
- d. -20°C (frozen)

4. When running External Controls on Sofia or Sofia 2, which control is run first if not selecting the skip (>>) icon?

- a. Negative Control
- b. Positive Control
- c. Either Negative or Positive Control

5. For proper test performance, which swab should be used to collect the sample and test as soon as possible?

- a. The swab supplied in the kit
- b. Sample type, swab, and storage conditions listed in the Package Insert - Sample Collection section
- c. Any type of swab
- d. Both A and B

6. How much processed sample solution is dispensed into the Sofia Test Cassette?

- a. 2 pipettes full, using 120 µL pipette (~240 µL)
- b. Entire amount of processed sample solution in Reagent Tube
- c. 4 drops
- d. 1 pipette full, using 120 µL pipette provided in the kit

7. How long is the swab left in the Reagent Tube?

- a. 30 seconds
- b. 1 minute
- c. 2 minutes
- d. 3 minutes

8. In the WALK AWAY Mode, how long does the Test Cassette sit on the counter prior to inserting into Sofia or Sofia 2?

- a. Cassette is placed immediately inside analyzer
- b. 5 minutes
- c. 15 minutes
- d. 10 minutes

9. In the READ NOW Mode, how long does the Test Cassette sit on the counter prior to inserting into Sofia or Sofia 2?*

- a. Cassette is placed immediately inside analyzer
- b. 5 minutes
- c. 15 minutes
- d. 10 minutes

*If trained on Strep A+ in addition to other tests, underline the Strep A+ answer and circle the answer that applies to the other tests.

10. Can a test line be seen on the Sofia Test Cassettes after development?

- a. Yes, you can see a pink test line
- b. No, Sofia and Sofia 2 detect the test line via fluorescent signal. The test line will not be visible to the user.



Select the assay(s) for which training was given:

SARS Antigen

Influenza A+B

Flu + SARS Antigen

RSV

Strep A+

Fill in or click the correct answer

This quiz is an educational tool intended to assist facilities in evaluating their operators' understanding of the assay(s) selected above. This quiz is not intended to be used as sole evidence of operator training or competency. Facilities are responsible for ensuring the quality of the testing performed by their operators. When testing controls or patient specimens, follow the current Package Insert instructions and/or Quick Reference instructions on quidel.com.

1. How often should the Calibration Check procedure be run?

- a. Every 14 days c. Anytime
 b. Every 30 days d. With every new kit lot

2. How often does the manufacturer recommend that external controls be run on the kit?

- a. Once every 20 tests c. Every first sample tested each day
 b. Once per each shipment of kits **d. Once per each untrained operator, once for each new shipment of kits and/or new lot of kits**

3. At what temperature do the kits need to be stored?

- a. 2°C to 8°C (refrigerated) c. 2°C to 30°C (room temperature or refrigerated)
 b. 15°C to 30°C (room temperature) d. -20°C (frozen)

4. When running External Controls on Sofia or Sofia 2, which control is run first if not selecting the skip (>>) icon?

- a. Negative Control c. Either Negative or Positive Control
 b. Positive Control

5. For proper test performance, which swab should be used to collect the sample and test as soon as possible?

- a. The swab supplied in the kit c. Any type of swab
 b. Sample type, swab, and storage conditions listed in the Package Insert - Sample Collection section **d. Both A and B**

6. How much processed sample solution is dispensed into the Sofia Test Cassette?

- a. 2 pipettes full, using 120 µL pipette (~240 µL) c. 4 drops
 b. Entire amount of processed sample solution in Reagent Tube **d. 1 pipette full, using 120 µL pipette provided in the kit**

7. How long is the swab left in the Reagent Tube?

- a. 30 seconds c. 2 minutes
 b. 1 minute d. 3 minutes

8. In the WALK AWAY Mode, how long does the Test Cassette sit on the counter prior to inserting into Sofia or Sofia 2?

- a. Cassette is placed immediately inside analyzer** c. 15 minutes
 b. 5 minutes d. 10 minutes

9. In the READ NOW Mode, how long does the Test Cassette sit on the counter prior to inserting into Sofia or Sofia 2?

- a. Cassette is placed immediately inside analyzer **c. 15 minutes (All other tests)**
 b. 5 minutes (Strep A+) d. 10 minutes

*If trained on Strep A+ in addition to other tests, underline the Strep A+ answer and circle the answer that applies to the other tests.

10. Can a test line be seen on the Sofia Test Cassettes after development?

- a. Yes, you can see a pink test line **b. No, Sofia and Sofia 2 detect the test line via fluorescent signal. The test line will not be visible to the user.**